



ENGLISH

Insert line for Periodontics

- INSERT #PLX1 Use: left-curved model for periodontal debridement on hard-to-access approximal surfaces.
- INSERT #PLX2 Use: right-curved model for periodontal debridement on hard-to-access approximal surfaces.
- INSERT #PLX3 Use: model for the irrigation and disinfection of periodontal pockets with antimicrobial solution.
- INSERT #PLX4 Use: left-curved model with ball-shaped tip for debridement in root furcations and concavities.
- INSERT #PLX5 Use: right-curved model with ball-shaped tip for debridement in root furcations and concavities.

Power setting (min. ÷ max.): 1 ÷ 7
Irrigation flow rate: medium ÷ high
Never direct the tip of the insert towards the tooth surface.

The Insert is compatible with the TITANUS®E (REF 400.00.OX and REF 410.00.OX) and EMS® piezo-electric Scalers. To remove or fix the Insert onto the scaler, use the supplied wrench. To remove or fix the Insert onto the scaler, use the special DYNO dynamometric wrench (REF 109.00).
 For use, cleaning and sterilization of the Insert, please refer to the scaler's operating instructions.

ITALIANO

Linea di inserti per Periodonzia

- INSERTO #PLX1 Uso: modello con curvatura a sinistra per curettaggio periodontale sulle superfici approssimali di difficile accesso.
- INSERTO #PLX2 Uso: modello con curvatura a destra per curettaggio periodontale sulle superfici approssimali di difficile accesso.
- INSERTO #PLX3 Uso: modello per l'irrigazione e la disinfezione di tasche periodontali con soluzione antimicrobica.
- INSERTO #PLX4 Uso: modello con curvatura a sinistra e con punta sferica per il curettaggio delle concavità e delle singole radici.
- INSERTO #PLX5 Uso: modello con curvatura a destra e con punta sferica per

il curettaggio delle concavità e delle singole radici.

Potenza (min. ÷ max.): 1 ÷ 7
Flusso di irrigazione: medio ÷ alto
Non dirigere mai la punta dell'inserto verso la superficie del dente.

L'inserto è compatibile con gli ablatori piezo-elettrici TITANUS®E (REF 400.00.OX e REF 410.00.OX) e EMS®. Per rimuovere o bloccare l'inserto sull'ablatore, usare la chiavetta in dotazione. Per rimuovere o bloccare l'inserto sull'ablatore usare la speciale chiavetta dinamometrica DYNO (REF 109.00). Per l'uso, la pulizia e la sterilizzazione dell'inserto, riferirsi alle istruzioni operative dell'ablatore.

REF 430.1n



0051

INSERT #PLXn
 INSERTO #PLXn



TeKne Dental s.r.l.

Via del Pescinale, 77
 50041 Calenzano (FI) - ITALY

TKD430.1n/Rev.09.2018

